

Herstellerbescheinigung

Manufacturer Certification



Hersteller: Manufacturer:	SIRONA Dental Systems GmbH
Adresse: Address:	Fabrikstraße 31 64625 Bensheim
Medizinprodukt: Medical device:	Autoklav DAC Professional, D 3472

Wir erklären die Übereinstimmung des DAC Professional mit der
EN 13060 Typ B
zur Dampfsterilisation aller verpackten oder unverpackten massiven Produkte, Hohlkörper des Typs A und poröser Produkte.

Nach der Richtlinie des Robert-Koch-Instituts „Infektionsprävention in der Zahnheilkunde – Anforderungen an die Hygiene“ von 2006 können mit dem DAC Professional Dentalinstrumente mit den Risikobewertungen

**Semikritisch A und B sowie
Kritisch A und B**

behandelt werden.

*We declare the conformity of DAC Professional with the
EN 13060 type B
for the steam sterilisation of all wrapped or non wrapped, solid, hollow load products type A and porous products.*

*According to the German RKI-Guideline (Richtlinie des Robert-Koch-Instituts) „Infektionsprävention in der Zahnheilkunde – Anforderungen an die Hygiene“ from 2006
it is possible to treat dental instruments with the DAC Professional in the risk evaluations
**semi critical A and B as well as
critical A und B.***

Bensheim, den: 28.8.06
Ort und Datum/Place and date

[Unterschrift/Signature]
[Name in Druckschrift/Printed Name] Mr. J. Siefert
Geschäftsbereichsleiter Instrumente/
Vice President Instruments Division

[Unterschrift/Signature]
[Name in Druckschrift/Printed Name] Mr. W. Müller
Approbations-Ingenieur/Approval-Engineer